




**LANIR-TEC LTD**  
18 Yitzhak Rabin St ;  
Modiin 7174986, Israel; Modiin 7174986  
T: 972-528474155 : F: 972-0737017209  
www.lanir-tec.com info@lanir-tec.com

## VSG301 for quick Analyzers and Monitoring systems tests

### VSG301 for Analyzers and Monitoring Systems tests





**LANIR-TEC LTD**  
18 Yitzhak Rabin St ;  
Modiin 7174986, Israel; Modiin 7174986  
T: 972-528474155 : F: 972-0737017209  
www.lanir-tec.com info@lanir-tec.com

### Quick Vibration instrumentation tests using VSG301 “live “ ISO 10816-3 Table

Connect all parts of the system under test:

- Laptop with Lanir-tec VSG301 -Vibration Signals Generator application
- VSG301 –LANIR-TEC Vibration Signals Generator
- TPI 9080 – Smart Vibration Analyzer


Run the VibrationSim.exe Application



**On the main menu – select the ISO 10816-3 button**




LANIR-TEC LTD  
 18 Yitzhak Rabin St ;  
 Modiin 7174986, Israel; Modiin 7174986  
 T: 972-528474155 : F: 972-0737017209  
 www.lanir-tec.com info@lanir-tec.com



Now please run any signal from the ISO 10816-3 table to test Analyzers or Monitoring systems.

“Green” signal level check. Signal level from 0.71 till 4.5 mm/s



“Yellow” signal level check Signal level from 4.7 till 7.1 mm/s



“Red” signal level check . Signal level from 7.3 till 20mm/s

